

Insulin Patch Pump System MTM-I equil 12:08 -7.8 Last BG 0.95 Last Bolus 1.6 **Current Basal** BASAL X D |

Insulin Patch Pump System User Manual

Table of Contents

1	Foreword	1
2	Brief Introduction	1
2.1	Indications	1
2.2	Patients	1
2.3	Contraindications	2
2.4	Before Use	2
2.5	How to Use This Guide	3
2.6	Getting Help	3
2.7	Prepare for Emergencies	4
2.8	Insulin Pump Precautions	4
2.9	Personal Diabetes Assistant	
	(PDA) Precautions	5
2.10	Blood Glucose Test Function	
	Precautions	5
2.11	Important Safety Information	6
3	Component Descriptions	8
3.1	Pump System Components	8
3.2	Accessories	8
4	Getting Started	11
4.1	PDA Setup	11

4.2	Charging the Batteries	15
4.3	Settings Wizard	17
4.4	Home Screen	24
4.5	Lock Screen	27
5	Basal Rate Adjustment	29

5.1 Understanding Basal Rate	29
5.2 Base Basal Rate	29
5.3 Editing Basal Programs (List Display)	29
5.4 Editing Basal Programs (Graph Display)	33
5.5 Activating a Basal Program	34
5.6 Temporary Basal Rate	35
5.7 Settings	36

6 Boluses	38
6.1 Understanding Boluses	38
6.2 Manual Boluses	38
6.3 Bolus Presets	39
6.4 Extended Boluses	41
6.5 Quick Bolus	44
6.6 Canceling a Bolus	46
6.7 Settings	47

7	Therapy Setup	49
7.1	Infusion Set Setup	49
7.2	Removing the Infusion Set	51
7.3	Filling the Reservoir	52
7.4	Assembling Reservoir to Pump	54
7.5	Disassembling the Reservoir	55
8	Replacing the Patch Pump	56
9	Replacing the Reservoir	64
9.1	Setup	64
10	The Integrated Blood Glucose Meter	70
10.	1 Operating Principle	70
10.2	2 Application	70
10.3	3 Blood Sampling	70
10.4	4 Lancet Removal	73
10.	5 Ejecting Test Strips	75
10.6	6 Testing Blood Glucose	75
10.7	7 Comparing Meter and Lab	
	Results	77
10.8	8 Quality Control Tests	78
10.9	9 Entering Your Glucose	

	49		Reading Manually	81
	49	10.10	Settings	82
	51	10.11	Blood Glucose Meter	
	52		Troubleshooting	84
р	54			
	55	11	Bolus Calculator	85
		11.1	Introduction	85
	56	11.2	Using the Bolus Calculator	86
		11.3	Settings	89
	64			
	64	12	History	91
		12.1	Viewing Your History	91
	70	12.2	Historical Averages	93
	10			
	70	13	General Settings	95
	70	13.1	Time and Date	95
	70	13.2	About the System	95
	73	13.3	Language	96
	75	13.4	Memory Card	96
	75	13.5	Display	96
		13.6	User Settings	97
	77	13.7	Bluetooth	97

14	Additional Features	100	17.10	Iranspo
1/1	Audio Playor	100	17.11	Storage
14.1		100	17.12	Other C
14.2	Auto-Off	100		
14.3	Food Database	100	18 18 1	<mark>Specif</mark> i Genera
15	Suspend/Resume	101	10.1	Dolivor
15.1	How to Suspend/Resume	101	18.3	Blood G
16	Alarms and Troubleshooting	103	10.4	
16.1	Patch Pump Alarms	105	18.5	
16.2	PDA Alarms	106	18.6	Infusior
16.3	Alarm System Delay	107	18.7	Occlusi
	<u>,</u>		(Maximu
17	Maintenance	108	18.8	Occlusi
17.1	Cleaning	108	18.9	Overdo
17.2	Avoid Extreme Temperatures	108	18.10	Electro
17.3	Avoid Immersion in Water	109		Compa
17.4	Test Strips	109	10	A m m m m
17.5	Control Solution	110	19	Append
17.6	X-Ray, MRI, and CT Scans	112	19.1	Symbol
17.7	Wireless Connection	112		
17.8	Precautions	112		
17.9	Disposal	113		

17.10	17.10 Transportation	113	
17.11	Storage	113	
17.12	Other Considerations	114	
18	Specifications	115	
18.1	General Specifications	115	
18.2	Delivery	116	
18.3	Blood Glucose Meter	116	
18.4	Bolus Calculator	117	
18.5 Bolus Delivery	117		
18.6	18.6 Infusion Precision	117	
18.7	Occlusion Detection		
	(Maximum Infusion Pressure)	118	
18.8	Occlusion Alarm Time	118 118	
18.9	Overdose/Underdose		
18.10) Electromagnetic		
	Compatibility	119	
19	Appendix	124	
19.1	Symbols	124	

1 Foreword

Thank you for choosing the Equil[™] Insulin Patch Pump system. We believe that Equil[™] will help you gain better control of your diabetes to be able to live a healthy and active life.

This user guide is designed to help you understand insulin pump therapy and the operation of your Equil[™] patch pump. We strongly recommend that you work closely with your healthcare professional to ensure you understand the pump functions and can begin therapy safely and effectively.

2 Introduction

2.1 Indications

This product is intended for subcutaneous (under the skin) delivery of insulin at set and variable rates for the management of diabetes mellitus in persons requiring insulin and for the quantitative measurement of glucose in fresh whole capillary blood (in vitro).

2.2 Patients

- This insulin pump system is suitable for patients with diabetes requiring shortterm or long-term insulin pump therapy.
- The blood glucose test function of the Personal Diabetes Assistant (PDA) is suitable for measuring the blood glucose level of whole blood samples that meet the following requirements:
- 1 Hematocrit range of 30% to 55%.
- 2 Triglyceride concentration does not exceed 33.9 mmol/L or cholesterol

concentration does not exceed 5.6 mmol/L. 3 Non-critically ill patients (such as those with severe dehydration or ketoacidosis etc.).

2.3 Contraindications

Patients who:

- Are unwilling to receive insulin treatment.
- Are unable to control their blood glucose, or unwilling to implant a cannula.
- Suffer from alcoholism, drug abuse, severe mental disorders (such as depression, schizophrenia).
- Suffer from allergies, including allergies to insulin and severe skin irritation.
- Are unconscious.
- Unable to understand or master insulin therapy concepts.
- Have severe hearing or vision impairment.
- Are elderly and living alone.
- Are children too young to administer diabetes therapy themselves.

2.4 Before Use

Before starting to use your Equil system, your healthcare professional should provide some important information regarding the management of your diabetes. If you have further questions regarding this information, please contact your healthcare professional for more guidance.

1 Basal Rate

Basal insulin is administered to maintain target blood glucose levels when not eating. You can configure up to three basal programs with Equil, which can help you adapt to different situations (for example, workdays, weekends, and sick days). Each basal program can be configured with up to 48 scheduled basal rate changes within one day. However, if you are new to using an insulin pump, you may decide to only use one or two basal rates during the day.

Active Insulin Time

The length of time that insulin

remains active and available in your body after a correction bolus. Quick acting U100 insulin should be used in the pump.

3 Target Blood Glucose Level

Insulin pump therapy requires a target blood glucose level. The purpose of an insulin pump is to keep the patient's blood glucose within a target blood glucose level range.

Insulin Sensitivity Factor

How much one unit of insulin can decrease blood glucose level. This number is used to calculate bolus injection amounts.

6 Carbohydrate Ratio

The number of grams of carbohydrates that can be covered by 1 unit of insulin.

2.5 How to Use This Guide

We recommend that you read this user guide thoroughly. Your healthcare professional can help you understand the usage in even greater detail. Please read this user guide in the correct order. In many cases subsequent sections refer to information detailed in previous sections.

🕑 Note

This user guide shows sample screens only. Your Personal Diabetes Assistant (PDA) screen may be slightly different.

2.6 Getting Help

This user guide describes the insulin pump system in great detail. However, you should still consult your healthcare professional for guidance. New users should consult a healthcare professional to help with first time setup and training of the insulin pump system.

Please seek help from your healthcare professional if you encounter problems. The manufacturer, Microtech Medical, and the local distributor, Homemed (Pty) Ltd, can provide technical support for the device, but cannot provide treatment advice for your medical condition.

2.7 Prepare for Emergencies

Patients with diabetes should always carry an emergency kit to quickly respond to any diabetes emergencies. Your emergency kit should include the following items:

- Blood glucose test strips
- Ketone test strips
- 1-2 sets of pump disposables (reservoirs, infusion sets, etc.)
- PDA and Pump battery chargers
- Alcohol prep swabs
- Backup pump battery
- Vial or cartridge of rapid acting U-100 insulin that is approved for your pump
- Fast- and long-acting insulin for manually injecting insulin should the pump malfunction
- Glucose tablets or another fast-acting source of carbohydrates
- Instructions from your healthcare professional regarding the amount of insulin to administer if the pump is interrupted, as well as your healthcare professional's contact phone number in case of emergency

2.8 Insulin Pump Precautions

The pump is used to administer insulin for people with diabetes. If it is used incorrectly, it could cause life-threatening situations.

- Before using the pump, please read the user guide carefully. Patients must be trained by a healthcare professional and should only use the pump after they have mastered the operation.
- Your healthcare professional should develop a customized insulin delivery protocol for you. Your healthcare professional will adjust the settings and observe the effectiveness, possibly monitoring your blood glucose four times a day until the therapy is stable.
- You should keep in frequent contact with your healthcare professional. Basic pump adjustments should be made under close supervision with a qualified healthcare professional.
- You should have sufficient knowledge of diabetes, and how to regulate your blood glucose level using insulin delivery and diet. You should understand the

effects of hyper- and hypoglycemia and how to prevent these conditions.

- If the pump fails to deliver your needed insulin, immediately stop using the system and consider using your emergency kit to supplement insulin. Contact your healthcare professional and/or local distributor, Homemed (Pty) Ltd for support.
- Please be sure to operate in strict accordance with this user manual, as failure to follow the instructions may result in unsafe practices. Microtech Medical and Homemed (Pty) Ltd is not liable for any related legal obligations arising from misuse.

This product can only be used to administer U-100 insulin.

Only MicroTech Medical disposable supplies should be used with the pump.

2.9 Personal Diabetes Assistant (PDA) Precautions

The PDA is the main interface for the entire system. Please note the following precautions:

- Do not let others handle your PDA, except your qualified healthcare professional.
- Please keep the battery charged .
- Do not drop or immerse in water as this could cause a malfunction.

2.10 Blood Glucose Test Function Precautions

- The blood glucose testing function is for in vitro diagnostic use only.
- The PDA is only compatible with MicroTech Exactive EQ blood glucose test strips.
- Use only fresh whole blood samples to test your blood glucose.
- The blood glucose test function is not intended for use in the diagnosis or screening of diabetes mellitus, nor for the use on newborns.
- Test results may be inaccurate if hematocrit levels fall outside the test strip hematocrit range of 30-55%.
- High levels of vitamin C or other reducing agents can lead to inaccurate results.

- The blood glucose measurement range is 1.1-33.3mmol/L.
- Triglycerides above 33.9 mmol/L and cholesterol above 5.6 mmol/L will lead to inaccurate test results.
- Results may be inaccurate when testing patients with abnormally low blood pressure or those who are in shock.
- For patients with impaired peripheral circulation, collection of capillary blood from the approved sample sites is not advised as the results might not be a true reflection of the physiological blood glucose level. It may apply under the following circumstances: severe dehydration as a result of diabetic ketoacidosis or due to stress hyperglycemia, hyperosmolar non-ketotic coma, shock, decompensated heart failure NYHA Class IV or peripheral arterial occlusive disease.
- Patients undergoing oxygen therapy may yield falsely low results.
- Discard the used test strips and lancets according to your local infection control procedures.

2.11 Important Safety Information

2.11.1 Do Not Immerse the Pump in Water

The patch pump is water resistant and splashproof (IPX4). Do not fully submerge the pump in water. If you plan to shower, bath, swim, or participate in other water activities, please suspend insulin delivery and remove the pump from the base. After you completed these activities, you may re-attach the pump to the base and continue pump therapy.

If the pump is accidentally dropped in water, use a soft clean towel to dry the pump as soon as possible. If you believe that water has entered the pump, or if you observe any other possible pump malfunction, please remove the pump from the base, check your blood glucose level and take precautions as necessary. Symptoms of high blood glucose include fatigue, excessive thirst and nausea. You should always contact your healthcare professional if you experience excessively high or low blood glucose levels, or if you have any questions about your care.

2.11.2 Static Electricity

The insulin pump system is resistant to normal levels of static electricity (ESD) but high levels of shock may cause a software reset which can interrupt insulin delivery.

ESD are more likely in situations where the relative humidity is very low, such as inside a heated building during the winter in areas where it is cold outside. If you suspect that your pump has failed, follow the instructions listed in Section 16: Alarms and Troubleshooting.



2 Pump Battery



8 Pump Battery Charger



🔨 Warning

Use only accessories and disposables that are manufactured by MicroTech Medical, or with the Equil brand name. Use of non-standard components may affect safety of the device.

4 PDA Charger



<u> Marning</u>

The included charger can be used in wall outlets rated at AC 110V-250V 50-60Hz. Connection to outlets that are outside this range can cause damage.

6 PDA Charging Cable



🙂 Note

Only use MicroTech's electrical components and accessories. Other accessories may lead to safety problems, including inaccurate insulin delivery. Microtech Medical and Homemed (Pty) Ltd cannot be held liable for problems arising from the use of third party accessories

Summary of components included with the Equil Insulin Patch Pump System:

- Patch Pump: MTM-1
- Personal Diabetes Assistant (PDA): MTM-2
- Cannula Inserter
- Pump Battery
- Pump Battery Charger
- PDA Charger
- PDA Charging Cable

Disposable components not included with this system, that need to be purchased separately:

- Equil Insulin Reservoir
- Equil Insulin Infusion Set (Cannula & Insulin Pump Base)
- Microtech Exactive EQ blood glucose test strips

4 Getting Started

4.1 PDA Setup

1 Remove the battery cover: Holding the PDA firmly with one hand, lift up the battery cover by inserting your fingernail into the slot shown in Figure 4.



Figure 4

2 Insert a microSD memory card: Insert memory card into the slot as shown in Figure 5.



Figure 5

🕑 Note

The microSD memory card must be inserted with the battery removed. A microSD card is not provided with your pump package, but you may provide your own. The microSD card is not needed for the primary functions of the pump system. However, secondary functions such as data transfer capability and MP3 player functions will not work correctly without a memory card inserted. **3** Insert the Battery: Install the battery in the orientation shown in Figure 6.



4 Replace the Battery Cover: Replace the battery cover to the position shown in Figure 7. The battery cover should be attached securely around the perimeter of the PDA, and you may hear a snap sound as the hooks engage.





Figure 6

🙂 Note

Use only batteries and chargers from MicroTech Medical. Use of third party accessories may cause unexpected errors and will void your warranty.

5 The PDA controls are shown in

Figure 8:





1 (Power Button)

Power On: Press and hold this key and the PDA will vibrate, begin the startup process, and enter the home screen after about 30 seconds.

Screen Off: If the PDA display is on, pressing the Power Button will switch the LCD display off and the PDA will enter standby mode.

🙂 Note

The PDA display will also switch off andenter standby mode after a preset time. See Section 13.6 for more information.

Screen On: When in standby, pressing the power button will switch the display on and present the lock screen.

Power Off: If the display is on, press and hold the power button to open a dialog box to confirm that you would like to switch the device off.

(Help Button)

If you need help with commands or functions of the PDA, you can press the

help button for more information.

🕑 Note

The help button functions on the Home, Basal, Bolus and History screens.

3 合 (Home Screen Button)

Press this to return to the home screen.

) 🖆 (Back Button)

Press this to return to the previous screen, or close a dialog box.

🕑 Note

Some functions, such as the change reservoir wizard, will not allow you to return to the home screen or return to a previous screen before completion.

▲ (Up Button)

This can be used to navigate the interface without using the touchscreen. Choose up to change selection.

🔻 (Down Button)

This can be used to navigate the interface

without using the touchscreen. Choose down to change selection.

🤊 🏌 (Enter (Select) Button)

This can be used to navigate the interface without the touchscreen. Choose the Enter button to select an action.

🕑 Note

The Enter button can also be used for special functions, as explained in subsequent sections.

- 8 Display
- 3.2" color display with touchscreen.
- 9 Charger/Data Port

Connect the PDA charger to this port to recharge the battery. This port can also be used to transfer data to a personal computer using a data cable.

Blood Glucose Test Strip Port Inserting a MicroTech Exactive EQ blood glucose test strip into this port will activate the blood glucose test menu. For more information on how to use the blood glucose meter, refer to Section 10. 1) 3.5mm Headphone Jack

Test Strip EjectorUsed to eject a blood glucose test strip.See Section 10.5 .

13 Speaker

4.2 Charging the Batteries

The patch pump and PDA batteries should be fully charged before use.

🕑 Note

Use only batteries and chargers from MicroTech Medical. Use of third party accessories may cause unexpected errors and will void your warranty.

4.2.1 PDA Charger

1 The PDA battery can only be charged while inside the PDA. The PDA may be on or off during the charging process, however, the battery will charge faster if the PDA is off. 2 Insert the small end of the cable into the PDA, and the large end of the cable into the PDA charger as shown in Figure 9.



Figure 9

3 Connect the PDA charger to a power outlet.

If the PDA is on, the battery icon will change to the charging icon. If the PDA is off, a charging animation will appear.

🕑 Note

If the PDA battery charger does not work, do not try to repair it. Please contact your local supplier for a replacement.

4.2.2 Patch Pump Battery Charger

1 The patch pump battery is inserted into the charger as shown in Figure 10. You should hear a click sound indicating that the battery is correctly inserted.



Figure 10

2 Insert the small end of the cable into the pump battery charger as shown in Figure 11. The large end of the cable should be inserted into the PDA charger.



Figure 11

3 Plug the PDA charger into a wall outlet, and the blue LED will turn on, indicating that the battery is charging. The indicator will flash if the charger is faulty.

4 When the battery is fully charged, the blue LED will turn off and you can disassemble the cable, pump battery charger, and PDA charger.

5 Remove the pump battery as shown in Figure 12: lift the battery up from the charger groove.

🕑 Note

If the battery is fully charged, do not leave it inside the charger while disconnected from a wall outlet. Doing so may slowly deplete the battery.





🕑 Note

A full battery should be stored properly inside an insulated package. Storing a battery with metal objects may cause a short circuit, resulting in a decrease in capacity or even battery damage.

🔨 Warning

Do not touch any metal part of the charger when plugged into a wall outlet as it is a potential shock hazard.

4.3 Settings Wizard

Switch the PDA on by pressing the power button. The first time you switch the PDA on, you will see the Settings Wizard which will guide you through basic setup options.

🕑 Note

When you are in the Settings Wizard, the Home button and Back button will not work until you finish the Wizard settings.

Time and Date Settings

After entering the Settings Wizard, you will be presented with the Time and Date setup screen shown in Figure 13.

Settings Wizard	<u>ا</u> ج
Welcome to the Equil Insu Management System!	ulin
Please set the time and date	
Date 2017-01-30	>
Time 00:30	>
24-Hour Format	~
1 / 5 Time & Date	ext



a) Choose the Date setting to open a dialog box (Figure 14) that can be used to input the correct date. Use the "+" and "-" buttons to select the date, choose OK to save and exit or Cancel to exit without saving.



b) Choose the Time setting to open a dialog box (Figure 15) that can be used to input the time. Use the "+" and "- " buttons to select the time, choose OK to save and exit or Cancel to exit without saving.



Figure 15

🕑 Note

If the 12 hour format is chosen, am and pm will be shown.



Figure 16

c) If 24-Hour Format is checked, the time and history data will be displayed in 24 hour format. Uncheck the box to use 12hour format.

Basic Basal Settings

The second screen in the Settings Wizard shows the basic basal settings (Figure 17).



a) Choose the Maximum Basal Rate setting to open a dialog box (Figure 18) that can be used to input the maximum basal rate. Use the "+" and "-" buttons to select the appropriate maximum rate. Choose "OK" to save and exit or "Cancel" to exit without saving.



Figure 18

🙂 Note

This feature is used to limit the maximum amount of basal delivery that can be administered by accident or maloperation.

b) Choose the Base Basal Rate setting to open a dialog box (Figure 19) that can be used to input your base basal rate. Use the "+" and "- " buttons to select the appropriate value. Choose OK to save and exit or Cancel to exit without saving.



Note

This value is the most frequently used basal rate, and all other rates will be calculated based on this base rate.

c) Choose the Temporary Basal Rate setting to open a dialog box (Figure 20) that will give you options.

Off: Temporary basal rates are disabled. The temporary basal rate controls will be greyed out.

U/hr: Temporary basal rates will be displayed in units of insulin delivered per hour.

%: Temporary basal rates will be displayed relative to your current basal rate.



Figure 20

After you have finished choosing the Basic Basal settings, choose Next to enter the third Settings Wizard screen, or Previous to enter the previous setting screen.

Basic Bolus Settings

The third screen in the Settings Wizard shows the Basic Bolus settings (Figure 21).



allows you to set the interval used when increasing or decreasing bolus amounts.



Figure 22

b) The Maximum Bolus setting allows you to set an upper limit on the amount of insulin given in a single bolus (Figure 23). Use the "+" and "- " buttons to select the value. Choose OK to save and exit or Cancel to exit without saving.



Figure 23

Note

This feature is used to limit the maximum amount of bolus delivery that can be administered by accident or maloperation.

c) Choose the Extended Bolus setting to open a dialog box (Figure 24) that will give vou options.

Off: Extended Bolus function is disabled. Extended Bolus controls will be greyed out.

U: Amount now will be shown in U (units of Insulin).

%: Amount now will be shown as a percentage of the entire bolus.



Figure 24

After you have finished choosing the Basic Bolus settings, choose Next to enter the fourth Settings Wizard screen, or Back to enter the previous setting screen.

Advanced Bolus Settings

The fourth screen in the Settings Wizard shows the Advanced Bolus Settings (Figure 25).





a) The Quick Bolus function allows the user to administer a bolus using only one button. This feature is switched off by default. For more detailed information about this feature, see Section 6.7
b) The Bolus Calculator can help the user decide on the size of a bolus by taking into account blood glucose levels and other parameters. This feature is switched off by default. For more detailed information about this feature, see Section 11.3

🕑 Note

These two features are relatively advanced and MicroTech recommends proper training and practice before using these features.

After you have finished choosing the Advanced Bolus settings, choose Next to enter the fifth Settings Wizard screen, or Previous to enter the previous setting screen.

6 Alert Options

The fifth screen in the Settings Wizard shows the Alert Options (Figure 26). If you

choose the Low Reservoir Alarm Threshold setting, a dialog box will open that allows you to choose when the device will give you an alert warning when the amount of insulin remaining in the reservoir is low and that youshould consider filling a new reservoir. Use the "+" and "-" buttons to select the appropriate value.



Figure 26

After you have completed the Settings Wizard, choose Finish to finalize your settings and return to the Home Screen. To continue modifying settings on previous screens, choose Previous.

Note

After you have completed the Settings Wizard, you can use all of the basic functions of the pump system. However, please read the additional sections of this user guide as they describe the features in greater detail.

Note

After you have completed the Settings Wizard for the first time, you may return to the Wizard by going to the Home Screen and choosing Settings and then Settings Wizard.

4.4 Home Screen

The Home Screen is shown in Figure 27. It is divided into three main sections: the Status Bar, Information Display Area, and Function Buttons



Status Bar:

The Status Bar contains icons that describe the status of the pump and PDA. The icons shown on the left are generally related to the pump (blue), and the icons on the right are generally related to the PDA (green). Each Icon in the Status Bar is described below:

- **Username:** Displays the name of the user.
- Insulin Pump Serial Number: Displays the serial number of the pump that is currently connected
- Alert icon: If an alert appears, this icon will appear on the status bar.
- Insulin Remaining: Shows how much insulin remains by varying degree; if PDA wireless link is poor it displays
 Insulin Comparison of the comparison o
- **Pump Battery Level:** Shows the pump's battery level status in varying degrees; if PDA wireless link is poor, it displays ? .
- • • • Wireless Link Status: Shows the wireless link strength between pump and PDA. It dispalys connection is totally lost.
- Ø Blood Glucose Test Reminder (BG Reminder): This icon will appear on the status bar if a Test Reminder has been programmed.

PDA Battery Level: Shows the PDA's battery level status in varying degree. If thePDA is charging, it displays $\boxed{}$.

Press the button in the middle of the Status Bar and it will enlarge as in Figure 28.



Figure 28

🕑 Note

If more than one BG Reminder has been programmed, only the next BG Reminder is shown in the enlarged Status Bar. If no BG Reminders are programmed, nothing is shown.

Information Display Area:

Normally the Information Display Area shows the three most important pieces of information regarding your condition: your last blood glucose level, the last bolus administered, and the current basal rate. During special functions (bolus delivery, temporary basal rates, and suspend mode) the Information Display Area will show the status of those functions until the actions are completed.

Function Buttons:

1 Bolus: Pressing this button takes you to the Bolus Screen where you can administer a normal or extended bolus.

2 Basal: Pressing this button takes you to the Basal Screen where you can adjust your basal programs or start a temporary

basal rate.

3 Actions: Pressing this button takes you to the Actions Screen where you can start a new reservoir, initiate a new pump, manually enter your blood glucose level, access the food library, or access the music player.

4 History: Takes you to therapy history and historical averages.

5 Settings: Press this to go to the Settings Menu.

6 Suspend/Resume: Pressing this button allows you to quickly suspend or resume insulin delivery.

Note

For more detailed information, please refer to subsequent sections of this user guide explaining the functions in greater detail.

4.5 Lock Screen

If you press the power button to switch the display off, or the display switches off

automatically after a period of non-use, the display can be switched on again by pressing the power button. The PDA will enter the lock screen as shown in Figure 29.

Information Display Area



Figure 29

The Lock Screen is divided into three main sections: Time and Date, Information Display Area, and the Unlock Area.

Time and Date Display Area

Displays the current time, day, date and year.

Information Display Area

Shows your last blood glucose level, the last bolus administered, and the current basal rate; or that status of a special function.

🕑 Note

The content of the Information Display Area in the Home Screen is identical.

Unlock Area

Unlock the device by swiping from left to right in this area.

🕑 Note

You can also unlock the PDA by simultaneously holding the Down and Enter buttons.

5 Basal Rate Adjustment

5.1 Understanding Basal Rate

Basal insulin is the pump's way to simulate the secretion of non-mealtime insulin. Even if a non-diabetic person fasts all day, there still is a small amount of insulin secreted. The pump's basal insulin is used to simulate the micro-secretions of insulin.

5.2 Base Basal Rate

We have already set up the base basal rate in the Settings Wizard. The base basal rate is the default basal rate for your basal programs. In the next sections you will learn to edit a basal program by modifying the base basal rate in different time periods.

5.3 Editing Basal Programs (List Display)

In the Home Screen, press the basal function button to enter the basal screen shown in Figure 30. You will see three Basal Program Tabs, a Basal Time Period List, and Basal Program Action buttons.

Basal Program Tabs Basal Time Period List



🕑 Note

When the PDA is in portrait orientation, the List Display is shown. If the PDA is rotated into landscape orientation, you will enter the Graph Display. However, if you are editing a particular time period, rotating the display will not change the Display until you have completed the entry.

When you first use the Basal Screen, no basal program is active. If you choose to activate one of the three basal programs, the previous basal rate will automatically stop.

Browsing Settings

You can look at Basal Program Settings by reviewing the Basal Time Period List. Swipe up or down in this area to scroll and see hidden list entries.

Reviewing Basal Programs

The PDA provides the ability to run one of three basal programs. You can review different basal programs by choosing one of the Basal Program Tabs.

Quick Basal Program Setup

The pump provides a Quick Basal Program Setup option. By entering your total daily basal insulin and choosing a 6 or 24 hour time period, the PDA can automatically create a Basal Program for you.

Select the appropriate Basal Program Tab and the choose Change Settings to edit the program (Figure 31).

Choose the top row and a window will pop-up for you to enter your total basal amount and which time period you would like (6 or 24 time periods). After entering these settings, the system will automatically assign Basal Time Periods to the Basal Program. Now you can further edit the Basal Time Periods according to your needs.

The 🗶 button in the top row will reset all basal time periods.

🕑 Note

The Quick Basal Program Setup is only an estimate of your daily basal needs. Consult your healthcare professional for more specific dosages.





4 Editing Basal Programs

a) Add new basal rate: The base basal rate was set in the Settings Wizard, so by default all 24 hours of the basal program are set to this delivery rate. To add a new basal time period, press the Change Settings button to enter the basal edit screen as shown in Figure 32.



Figure 32
Press the S button to add a new basal time period (Figure 33). After editing is complete (Figure 34), the basal rate time will automatically be divided into multiple time periods.







Figure 34

b) Editing a basal time period: You can edit a basal time period entry by pressing the right icon.
c) Deleting a basal time period: You can delete a basal time period entry by pressing the right icon. The time period will revert to the base basal rate.

🕑 Note

After deleting a basal time period, the PDA may merge adjacent basal time periods if the basal rates are the same.

Changing the Basal Program Name

Press the Change Basal Program Name button to open the Change Name dialog box (Figure 35). Change to the desired name.

Chang	e Name
[]	
ок	Cancel

Figure 35

6 Saving Settings

After you have completed editing the basal program, press Save Settings to save and exit the Basal Edit screen.

🕑 Note

If you have edited a basal program that is currently running on the pump, a dialog box will open asking if you would like to send the updated program to the pump. Choose OK to update and save or Cancel to return to the Basal Edit screen.

🕑 Note

If you press the Back button(⇐), while in the Basal Edit screen, a dialog box will open to confirm that you would like to exit. Choose OK to discard your changes. Choose Cancel to continue editing.

5.4 Editing Basal Programs (Graph Display)

From the Home Screen, press the Basal Function Button to enter the Basal Screen, then choose the correct Basal Program tab and then rotate the PDA to landscape orientation. This will activate the Basal Program graph display as shown in Figure 36.

🕑 Note

Before editing the graph display, please review the functions of the list display to better understand the graph display functions.



Figure 36

Reviewing Basal Programs

The vertical axis of the chart represents the basal rate while the horizontal axis represents the time, which shows 24 hours of the day.

2 Add a Basal Time Period

a) In the graph display, press and hold on any area of the graph to enter the edit mode as in Figure 37. The highlighted area is the period to be edited.





b) By pressing the **(**) buttons, or the touchscreen, you can choose a different

basal time period.Press 💉 to pop up the Basal Rate Setting window as shown in Figure 34.

c) When finished, press 📃 to save and then 📉 to exit the editor mode.

🕑 Note

You cannot switch between Basal Programs in Graph Display mode. You must return to List Display to change Basal Program.

🙂 Note

If you have edited a basal program that is currently running on the pump, a dialog box will open asking if you would like to send the updated program to the pump. Choose OK to update and save or Cancel to return to the Basal Edit screen.

5.5 Activating a Basal Program

From the Home Screen, press the Basal Function Button to enter the Basal Screen shown in Figure 31. Select the correct basal program by choosing the correct tab. Press the "Activate <Program Name>" button, and confirm using the dialog box.

🕑 Note

After completing the Settings Wizard for the first time, you must start basal delivery by activating a basal program. Thereafter, activating a new basal program will supersede any programs that were previously running.

🕑 Note

When a basal program is running, the Activate Program button inside the running program's tab will change to Temporary Basal Rate (see Section 5.6).

5.6 Temporary Basal Rate

Normally, your basal rate does not require frequent changes, but in certain circumstances you may want to temporarily administer a different basal rate to avoid low or high blood glucose levels.

1 Your basal insulin requirements might change during periods of exercise, sick days and travelling. Climate changes, hormonal changes, menstruation, medication, alcohol and certain foods might also require an adjustment to the basal insulin rate. Your healthcare professional will be able to advise you on when and how to set a temporary basal rate.

🕑 Note

A temporary basal rate should be determined under guidance from your healthcare professional.

From the Home Screen, press the basal function button to enter the Basal Screen, make sure that the Current Basal Program tab is open, and press the Temporary Basal Rate button to bring up a temporary basal rate settings dialog box (Figure 38).

🕑 Note

The Rate can either be set in U/hr or by % of the current basal rate. For more info about settings, see Section 5.7



Enter your temporary basal rate and the amount of time. Press the Start button and the PDA will send the command to the pump to administer a temporary basal rate. When the command is completed, the PDA will automatically return to the Home Screen where the Information Display Area will show the current status of your temporary basal rate (Figure 39).



🕑 Note

By default, the Temporary Basal Rate function is off. If the button is greyed out, activate the function using the settings menu (Section 5.7)

5.7 Settings

From the Home Screen, choose "Settings – Insulin Delivery Settings" as shown in Figure 40 to change delivery settings.



1 Maximum Basal Rate: Used to limit the maximum basal rate that can be set in the Basal Menu

🕑 Note

This feature is used to limit the maximum amount of bolus delivery that can be administered by accident or maloperation.

2 Base Basal Rate: Used to enter the Base Basal Rate (Section 5.2)

🕑 Note

Setting this value to the most used basal rate will simplify creating Basal Programs.

3 Temporary Basal Rate: Choose to adjust settings for temporary basal rates.

a) **Off:** Switch off the Temporary Basal Rate function. The corresponding Temporary Basal Rate button will be greyed out.

b) **U/hr:** Temporary Basal Rate will be displayed in U/hr.

c) **%:** Temporary Basal Rate will be displayed as a % of the current basal rate.

6 Boluses

6.1 Understanding Boluses

People with diabetes have elevated blood glucose levels after meals, requiring bolus doses of insulin to lower blood glucose levels. A bolus refers to the amount of insulin that is quickly infused before a meal.

Carbohydrates, proteins and fats all affect your blood glucose level differently (Figure 41). There are two types of boluses that can be administered to address different types of food intake.





6.2 Manual Boluses

Manual boluses are used to offset high blood glucose levels.

From the Home Screen, choose the Bolus button to enter the Bolus Screen (Figure 42).



Figure 42

Choose the Bolus button in the upper left corner to open the dialog box (Figure 43). Choose your bolus amount and press OK.





If the Bolus amount is not 0, the Start button will be green, and you can choose the Start button to begin to administer the bolus. A window will open for you to confirm that you want to send the bolus. (Your last bolus amount and time will be shown as a reminder).

While the pump is delivering, the PDA will revert to the Home Screen and display the bolus status (Figure 44).



6.3 Bolus Presets

The Bolus Screen contains three programmable preset boluses (Figure 45).





You can set up each of the three preset boluses (they are empty by default).

1 Add Preset: If a preset is empty, you may press anywhere on the button to open the Preset dialog box (Figure 46). Edit the parameters to create a new Preset.





2 Using a Preset: If a Preset is already programmed, press the left area of the button and the value will appear in the bolus amount to the far left. A small arrow will appear that links the preset that has been chosen to the bolus amount (Figure 47). You can administer the bolus now by pressing the Start button as in Section 6.2.

Bolus	17:24	
Bolus	Lunch 🖍	
4.0	Preset 2 💉	
U	Preset 3 💉	
Extend Bolus		
Start		



3 Edit Preset: Press the *reset* icon to open the Preset settings dialog box. Change the settings, and press OK to save.

🕑 Note

All three presets can be administered in any order, but only one can be administered at a time.

6.4 Extended Boluses

6.4.1 Extended Bolus with Amount Now

An Extended Bolus with Amount Now can be useful when you eat meals with both rapidly and slowly absorbed carbohydrates including meals with a high fat content such as some fast foods. An Extended Bolus With Amount Now is a bolus that is administered in two stages – an amount immediately, and the remainder over an extended period of time.

🕑 Note

This feature requires a detailed understanding of ones carbohydrate absorption rate. Please consult your healthcare professional before using this feature. From the Home Screen, choose the Bolus button to enter the Bolus Screen and then enter a bolus amount (Section 6.2). Press Extend Bolus to open the Extend Bolus window (Figure 48).

🕑 Note

By default, the Extend Bolus function is off. If the button is greyed out, be sure to activate the function using the settings menu (Section 6.7).



Figure 48

Amount Now: The bolus amount to be administered immediately Extend amount: Total Bolus – Amount Now Extend Time: The time period over which the extended amount will be administered.

🙂 Note

The Amount Now can be set in absolute units (U) or relative to the total bolus amount (%). See Section 6.7 to change this setting.

Press OK to accept this setting and go back to the Bolus Screen. The Bolus Screen will now display the Amount Now and Extend Time (Figure 49).



Choose the Start button to begin to administer the bolus. A window will open for you to confirm that you want to send the bolus. (Your last bolus amount and time will be shown as a reminder), choose OK to confirm and the PDA will send a command to the pump to start delivery. While the pump is delivering, the PDA will revert to the Home Screen and display the bolus status (Figure 50).



6.4.2 Extended Bolus Without Amount Now

An Extended Bolus Without Amount Now can be useful when you have eaten a meal over a prolonged period of time with extended snacking, have delayed digestion due to gastroparesis or for meals high in fat. The Extended Bolus Without Amount Now can be administered over a time period between 30 minutes and 8 hours.

Before using this feature, please read Section 6.4.1. To administer an Extended Bolus Without Amount Now, set the Amount Now to 0 (zero). No immediate bolus will be administered, and the total bolus amount will be administered over the time period specified.

6.5 Quick Bolus

The Equil Insulin Patch Pump System has a Quick Bolus feature that allows the user to administer a bolus without using the touchscreen or looking at the PDA Display. This feature can be handy if you forget your PDA, or would like to administer a bolus discretely.

🕑 Note

By default, the Quick Bolus function is off. To enable, be sure to activate the function using the settings menu (Section 6.7)

6.5.1 Quick Bolus Using the PDA

1 While in the Lock Screen, press and hold the Enter (Select) button for about 3 seconds. The PDA will give an audible alert and enter the Quick Bolus screen (Figure 51).





2 The next time you press the Enter (Select) button the bolus size will increase by a predetermined increment. This increment can be set in the settings menu (Section 6.7). Also, every time you press the Enter (Select) button, you will hear a beep from the PDA to confirm that you have pressed the button. Each time you press the button, the tone will ascend in pitch (up to 5, and then repeating) to help you remember the number of times you have pressed the button.

🕑 Note

The maximum number of increments for a quick bolus is 20.

🕑 Note

If there is no activity for 10 seconds, the PDA will revert to the Lock Screen.

3 After you have set the correct bolus amount by pressing the Enter (Select) button the appropriate number of times, wait for 3 seconds and the PDA will repeat the audible tones back to you to confirm the correct number of bolus increments. If the number of tones is correct, press the Enter (Select) button one last time, and the bolus will be administered.

🕑 Note

If you would like to cancel, simply press the Back button or wait 10 seconds for the system to revert to the lock screen

When the pump receives the bolus correctly, you will hear a confirmation tone and the bolus will start. The PDA will automatically return to the main screen, displaying the bolus and time remaining.

6.5.2 Quick Bolus Using the Pump

Administering a Quick Bolus on the pump is similar to administering a Quick Bolus on the PDA.

1 Press and hold the Bolus button on the patch pump. The pump will give an audible alert which indicates that you can enter a quick bolus amount.

2 With each subsequent button press, the bolus size will increase by a predetermined increment. This increment can be set

in the settings menu (Section 6.7). Also, every time you press the Enter (Select) button, you will hear a beep from the pump to confirm that you have pressed the button. Each time you press the button, the tone will ascend in pitch (up to 5, and then repeating) to help you remember the number of times you have pressed the button.

🕑 Note

The maximum number of increments for a Quick Bolus is 20.

Note

If there is no activity for 10 seconds, the pump exits the Quick Bolus mode.

3 After you have set the correct bolus amount, wait for 3 seconds and the pump will repeat the audible tones back to you to confirm the correct number of bolus increments. If the number of tones is correct, press the pump button one last time, and the bolus will be administered.

🙂 Note

If you would like to cancel, simply wait 10 seconds for the system to exit the Quick Bolus mode.

6.6 Canceling a Bolus

A bolus can be stopped during delivery. This can be helpful, for example, if you notice that the bolus size is incorrect, or if your meal plans change.

When the bolus is being administered, the Home Screen is displayed and bolus information is shown in the Information Display Area.

Press the Cancel button and a confirmation window will appear. Press OK to confirm. You can also cancel a bolus by pressing and holding the Quick Bolus button on the pump or press and hold the "Bolus" button on the pump to cancel.

🙂 Note

Cancelling a bolus delivery will cancel the total bolus (Amount Now and Extended amount).

6.7 Settings

From the Home Screen, the Bolus settings can be found at "Settings - Bolus Settings" as shown in Figure 52.

Insulin Delivery Settings	
Maximum Basal Rate 1.5U/hr	>
Base Basal Rate 0.5U/hr	
Temporary Basal Rate OFF	
Bolus Rate 3.0U/min	
Bolus Increment 0.1U	
Maximum Bolus 10U	
Extended Bolus OFF	



1 Bolus Rate: The normal bolus delivery rate is 3.0U/min. Use this menu to decrease the bolus rate to 1.5U/min.

2 Bolus Increment Setting: This is the amount that the requested bolus changes when pressing the "+" or "-" button.

3 Maximum Bolus Setting: This is the maximum bolus that the user can choose in the bolus menu.

🕑 Note

This feature is used to reduce the possibility of overdosage due to error. Try to set this value as the maximum bolus that you would normally give. Setting this value too high increases the risk of overdosage by maloperation.

4 Extended Bolus Settings:

a) **OFF:** Extended bolus function is off, and corresponding button will be greyed out (default)

b) U: The Amount Now will be displayed in Units of insulin (U).

c) %: The Amount Now will be displayed

as a % of the total bolus amount.

5 Quick Bolus Settings: Check the checkbox next to Quick Bolus to enable the Quick Bolus function. Additional Quick Bolus settings will be activated.

a) **Quick Bolus Increment:** Sets the amount that the bolusincreases with every Quick Bolus button press.

7 Therapy Setup

7.1 Infusion Set Setup

1 Firstly wash your hands. Then use an alcohol wipe to clean the infusion site area, removing any oils/dirt. Allow it to air dry. Remove the pump base from its packaging and place it on one of the suggested areas shown in Figure 53 (abdomen, upper arm, thigh, lower back.) Avoid sites that may rub against other objects like belts, waistbands or tight clothing. Also make sure that your infusion site is at least 2-3cm away from your navel. Do not use the abdomen if you are in your 3rd trimester of pregnancy.



2 Open the cannula pack as indicated on the packaging. Push the cannula assembly into the cannula inserter as shown in Figure 54 until the cannula assembly is in the cocked position and you hear two

"click" sounds. You should see that the cannula assembly is connected firmly to the cannula inserter and the inserter is held firmly in the cocked position.

🕑 Note

Infusion sets are comprised of a cannula pack and base. Be sure to use only infusion accessories from MicroTech such as HRN-S-60, HRN-S-90, HRN-H-45, or HRN-H-65.







3 Remove the protective sleeve as shown in Figure 55.





4 Hold the cannula inserter and align the front end with the bottom plate of the pump base as shown in Figure 56. Press down, until you hear a "click" securing it in place. Press the release button on both sides of the cannula inserter at the same time, then the cannula will be inserted into the base and infusion site.



When pressing the release button, the other hand must be pressed against the bottom of the cannula inserter and the baseplate cannot be lifted. **5** Release the inserter from the pump base by pressing the button on the lower end of the cannula inserter as shown in Figure 57. If the inserted needle or needle base remains in the pump base, carefully remove it from the infusion site. Place the protective sleeve back onto the cannula assembly.



Figure 57

6 If the needle remains in the inserter, place the protective sleeve back onto the needle hub. Firmly push the top of the inserter to eject the needle hub (Figure 58) and dispose of properly.

🕑 Note

Dispose of the insertion needle properly in a medical sharps container.



Figure 58

7.2 Removing the Infusion Set

Peel away one end of the tape from the skin as shown in Figure 59 and continue until the entire infusion set has been removed.



Figure 59

7.3 Filling the Reservoir

1 Remove the new reservoir from the packaging.

2 Use an alcohol wipe to clean the insulin vial or cartridge and then attach it to the fill adapter as in Figure 60.





3 With the insulin vial or cartridge on top, slowly pull the pullrod and draw insulin into the reservoir as in Figure 61.





4 Tap the side of the reservoir to encourage any air bubbles to rise to the top of the reservoir (Figure 62).



Figure 62

5 Push the pullrod slowly to eject the bubbles back into the vial or cartridge, and then slowly pull to draw more insulin into the reservoir. Repeat until the reservoir has enough insulin and there are no visible air bubbles (Figure 63).



Figure 63

6 Remove the insulin vial or cartridge from the fill adapter (Figure 64).



Figure 64

7 Pull the two release tabs (shown in Figure 65) away from the reservoir to release the fill adapter.



Figure 65

8 Remove the pullrod by unscrewing it counterclockwise (Figure 66)



7.4 Assembling Reservoir to Pump

Figure 67 illustrates the sequence to attach the reservoir and battery to the pump. Always install a fully charged battery with a newly filled reservoir.







Figure 67

Be sure that the battery is inserted in the orientation shown in Figure 68.



Figure 68

7.5 Disassembling the Reservoir

Holding the pump and reservoir in the orientation shown in Figure 69, bend the assembly along the seam line as shown until the reservoir separates from the pump. Be sure to recharge the battery immediately to make sure that a fresh battery is always installed with a new reservoir. Dispose of the used reservoir in a medical sharps container.

🕑 Note

Pay close attention to the orientation of the pump during disassembly. Incorrect disassembly may cause damage.



Figure 69

8 Replacing the Patch Pump

λ Warning

Before continuing, please make sure you are familiar with frequently used pump setup operations such as installing/removing an infusion set, filling a new reservoir, and assembling/ disassembling the reservoir (Section 7).

This section explains how to connect a new patch pump to the PDA or replace an old one.

1 From the Home Screen, choose the Actions button and then the New Pump button to enter the New Pump Wizard

a) If this is your first time using the wizard to connect your pump, the wizard will start with Figure 70. Apply a new infusion set (Section 7.1), assemble a filled reservoir and a charged battery (Section 7.3 and 7.4) to the pump as shown in Figure 70. Press Continue to go to the next page.







Figure 70

b) If this is not your first time using the wizard to connect a pump, the wizard will start with Figure 71, as displayed on the next page. Remove the pump and reservoir from the pump base as shown in Figure 71. Remove the old infusion set (Section 7.2). apply a new one (Section 7.1) and press Continue. The next page will prompt you to deactivate your old pump (Figure 72). Press Continue to disconnect the old pump from the PDA and wait for the pump to rewind. Separate the used reservoir from the pump (Section 7.5), dispose of the old reservoir and then assemble a filled reservoir and a charged battery to the pump as shown in Figure 70 on previous page. Press Continue to go to the next page.





Be sure to use a fully charged battery. A low battery may require you to change reservoirs more frequently, resulting in wasted insulin.









2 Enter the New Pump Serial Number in the space provided (Figure 73) using the onscreen keyboard. Press the Back button to exit the keyboard and press Continue. The PDA will attempt to connect to the new pump. After the new pump has been activated, the PDA will show a confirmation screen (Figure 74).





Figure 73

3 From the confirmation screen (Figure 74), press the Continue button.



4 Holding the pump in the orientation shown in Figure 76, press the Prime Reservoir button (Figure 75). The plunger will begin to move slowly. Continue to press the button until you see a drop of insulin on the needle tip (Figure 76). Now press Continue.



Needle Tip



5 Connect the pump to the infusion set as shown in Figure 77 and press Continue.









6 Now choose whether or not to prime the cannula (Figure 78). Upon completion, the pump will begin to deliver.

🕑 Note

Skip only if the infusion set was not replaced (and thus the cannula doesn't need to be filled).





9 Replacing the Reservoir

9.1 Setup

🚹 Warning

Before continuing, please make sure you are familiar with frequently used pump setup operations such as installing/removing an infusion set, filling a new reservoir, and assembling/ disassembling the reservoir (Section 7).

If the insulin reservoir connected to your pump is empty, you should replace the reservoir.

1 From the Home Screen, choose the Actions button and then the New Reservoir button to enter the New Reservoir Wizard. Remove the pump and reservoir from the infusion set as shown in Figure 79. Press Continue, the pump will rewind, and the PDA will display the next page of the wizard. Before proceeding to the next step, you should remove and replace the infusion set (Section 7.1-7.2) and prepare a new insulin reservoir (Section 7.3).









2 Separate the used reservoir from the pump (Section 7.5), dispose of the old reservoir, and then assemble a filled reservoir and a charged battery to the pump as shown in Figure 80. Press Continue to go to the next page.

🙂 Note

Be sure to use a fully charged battery. A low battery may require you to change reservoirs more frequently, resulting in wasted insulin.







Figure 80

3 Holding the pump in the orientation shown in Figure 82, press the Prime Reservoir button (Figure 81). The plunger will begin to move slowly. Continue to press the button until you see a drop of insulin on the needle tip (Figure 82). The background of the "Next" button will be enabled when it changes from grey to green. Press Continue to go to the next step.




4 Connect the pump to the infusion set as shown in Figure 83 and press Continue.









5 Now choose whether or not to prime the cannula (Figure 84). Upon completion, the pump will begin to deliver.

🕑 Note

Skip only if the infusion set was not replaced (and thus cannula doesn't need to be filled).



Figure 84

10 The Integrated Blood Glucose Meter

10.1 Operating Principle

The integrated blood glucose test meter uses an electrochemical biosensor containing glucose oxidase and detects β -d glucose in a blood sample. When the blood sample touches the edge of the test strip, the strip automatically absorbs it and reacts with the chemical reagent in the reaction area. The chemical reaction causes a change in current, which is measured to yield the glucose concentration.

10.2 Application

It is used to determine blood glucose concentrations in fresh capillary whole blood. It can be used for both self-testing and professional use. Test results can be used to manage blood glucose levels, but should not be used for the diagnosis of diabetes.

10.3 Blood Sampling

Before testing, become familiar with how to collect a blood sample and choose a clean and dry place to conduct the test.

Important Tips

Prior to testing, either use an alcohol wipe or soapy water to disinfect the sampling site. Use war water to increase blood flow if necessary. Dry your hands and the sampling site, ensuring that there is no soap residue remaining.



10.3.1 Fingertip Testing

Adjust the penetration depth of the lancing device to reduce the level of discomfort.

1 Remove the lancing device cap. Insert the lancet into the lancet holder until it comes to a complete stop as indicated in Figure 86.





Figure 86

2 Twist off the safety cap from the lancet, (Figure 87) and save the safety cap for lancet disposal.





3 Carefully install the lancing device cap onto the lancing device, avoid touching the lancet needle tip (Figure 88).





4 Adjust the puncture depth by rotating the depth adjustor (Figure 89). The lancing device has 5 puncture depth settings. To reduce discomfort, choose the lowest setting that still produces an adequeate blood sample.

Depth Adju	ıstment
1 and 2:	for delicate skin
3:	for normal skin
4 and 5:	for thick or calloused skin

🕑 Note

Greater pressure between the lancing device against the finger will also increase the puncture depth.





5 Pull back the cocking barrel (Figure 90) until you hear a click. Now the lancing device is loaded and ready to draw blood.



Figure 90

6 Before taking a blood sample, wash your hands or use an alcohol wipe to clean the area. Washing your hands in hot water increases blood circulation. You can also massage from wrist to finger to promote better blood circulation (Figure 91).



Figure 91

7 Holding the lancing device against the side of the finger to be lanced, press the release button and then put down the lancing device. Massage forward slowly from the base of your finger to the tip to increase the sample size (Figure 92).





🕑 Note

To reduce pain, lance on the sides of the fingertips, where there are less nerve endings. Rotate finger blood sample sites as much as possible to accelerate wound healing and decrease callouses.

10.4 Lancet Removal

1 Unscrew the lancing device cap. Firmly push the needle into the safety cap as shown in Figure 93.



Figure 93

2 Pull out the lancet from the lancet holder and dispose of the lancet in a sharps container. (Figure 94)



Figure 94

Lancet Precautions:

- Do not use a lancet if the safety cap is loose or missing.
- Do not use a lancet if the needle is bent.
- Use caution whenever a lancet needle is exposed.
- Do not share lancets with other people.
- To avoid cross contamination, always use a new sterilized lancet. Do not reuse lancets.
- Avoid contaminating lancets with hand lotion, detergents, oil, and other debris.

Reminder:

- Lancing devices and lancets should not be shared. Each person should have his own lancing device and lancets.
- Clean your lancing device before and after use with an alcohol wipe or a disinfectant wipe. Be sure to clean the part of the device that touches the finger. Do not immerse the lancing device in water.
- Control excess bleeding and disinfect your wound after use.

10.5 Ejecting Test Strips

The back side of the PDA contains a test strip ejector. You can easily eject the test strip by sliding the ejector button as shown in Figure 95.





10.6 Testing Blood Glucose

If the display is on, insert an Exactive EQ test strip and the blood glucose meter screen will appear (Figure 96).



Figure 96

After the strip has been inserted, apply blood as shown on the screen. When

enough blood has been applied, the screen will count down 5 seconds and then display the test result as in Figure 97.



1 Test Result Area: Displays the blood glucose test result with the correct unit and time/date.

The scale above the reading shows whether or not the result falls within the target range. If the result is outside of the range, the bar will be yellow.

2 User Actions Area: Contains blood glucose result tags, invalid measurement selection and Bolus Calculator button.

Result Tags: Use this area to tag your results before/after exercise, or before/ after meal. These tags will help sort your results into different categories when calculating averages.

Invalid Measurement Box: if checked, the measurement will be recorded in history, but will not be used to calculate averages.

Bolus Calculator: If the Bolus Calculator is enabled in the Settings Menu, the button will turn green after a blood glucose reading has been recorded.

After the test has been completed, slide the test strip ejector to pop out the test strip. Ejecting the test strip or pressing the back button will return the display to your last screen

🕑 Note

Your blood glucose level is automatically saved to history upon leaving the BG meter function.

As shown in Figure 98, apply the blood sample to the edge of the strip until the test window is full.

The PDA screen will begin to count down when there is adequate blood. If the test window is not full, you may add additional blood within three seconds. If not enough blood has been applied, an error will be displayed. Please discard and use a new test strip. If you see that the test window is not full, but the countdown begins anyway, please discard and use a new test strip.



Figure 98

10.7 Comparing Meter and Lab Results

Your PDA's blood glucose meter and laboratory equipment both report glucose concentrations in the serum or plasma component of your blood. However variations between the two are normal, and your meter results and laboratory results may be slightly different.

To ensure a reasonable comparison between your meter and laboratory results, please follow these guidelines:

1 Make sure the PDA's meter is working properly.

2 Comparisons will be more accurate if you do not eat for at least four hours (preferably eight hours) before testing.

3 Bring your PDA, test strips, and control solution to the lab.

4 Ensure that the time between tests with your PDA and the laboratory is within 15 minutes.

5 Wash and dry your hands before obtaining a blood sample.

6 Make sure you closely follow the instructions in this manual.

Test results may show small deviations, this may due to the following reasons:

Blood oxygen and red blood cell count vary from person to person, and even within the same person. The glucose meter tests blood glucose concentrations for the widest range possible. If the user's blood indexes fall within the middle of the range, the result will be ideal. Otherwise, there will be come small deviations. (The deviations should be within the range allowed by local goverment.)

10.8 Quality Control Tests

Control solution is a glucose solution of known concentration that is used to confirm that the PDA's meter and test strips are working properly. Normally, you should use Control Solution 1 and Control Solution 2 should only be used for secondary testing. Control Solution must be purchased separately. Please use control solution to perform quality control tests which can verify that the PDA's meter is working properly.

You should perform a control solution test if you suspect that the meter or test strips are not working properly, if you suspect that your test results are inaccurate or inconsistent with the way you feel, or if you suspect the meter is damaged:

1 If the display is on, insert a test strip and the blood glucose meter screen will appear. Check the control solution check box to indicate that you are doing a quality control test. The PDA will display an animation as in Figure 99. Shake the control solution bottle, gently squeeze out the control solution, discard the first drop and drop the second drop onto a clean nonabsorbent surface. Now touch the second drop to the sample area of the test strip. Do not let the bottle come into contact with the test strip.







2 When enough control solution has been applied, the screen will count down 5 seconds and then display the test result as in Figure 100. The result is displayed in the top half of the screen. If the result falls within the range printed on the test strip package (typically CTRL1) the device is working properly.

🕑 Note

If a big bubble forms, wipe with clean cotton wool, and then do the following: If the test window is not full, do not add more solution. Discard and try again with a new strip.

I1:27 * • Control Solution Test 2 13 14 15 16 17 18 19 20 21
- 16.8 +
mmol/L
📕 Before Exercise 🦷 Before Meal
After Exercise After Meal
✓ Invalid Measurement
Bolus Calculator



3 After the test has been completed, slide the test strip ejector to pop out the test strip.

🕑 Note

The control solution test results will not be stored in your history and the test result tags will be greyed out. If the control solution results are outside of the reference range:

- Confirm you are matching the correct range. Control Solution 1 results should be matched to the CTRL1 range printed on the test strip vial (or foil pouch).
- Check the expiration date of the test strip and control solution. Make sure that the packages have not been opened for more than 6 months. Discard any expired test strips and control solution.
- Confirm that you are testing within the correct temperature range (15-30 °C).
- Make sure that the test strip vial and control solution bottle have been tightly closed.
- Make sure that you are using the correct brand of control solution.
- Make sure you are following the user guide instructions properly.

After checking all of the conditions above, repeat the quality control test with a new test strip.

If the quality control test results are still

outside of the range printed on the test vial (or foil pouch), there may be a problem with your meter. Please contact your supplier.

Control Solution 1 is sufficient for most self testing needs. If you think the PDA or strips may not be working correctly, you may also want to do a level 2 test. Repeat the above steps using Control Solution 2, see if the results fall within the CTRL2 (Control fluid 2) range printed on the label. For confirmation of results, Control Solution 1 tests should fall within the CTRL1 range, and Control Solution 2 tests should fall within the CTRL2 range.

10.9 Entering Your Glucose Reading Manually

You can also enter your blood glucose reading manually. This is especially useful if you use a separate blood glucose meter. The entry will be stored in your history and can be used with the bolus calculator. From the Home Screen, choose the Actions button and then the Enter BG Reading Manually button to enter the manual entry screen (Figure 101). Using the

"+" and "-" buttons, enter your blood glucose reading, select the appropriate tag (Before/After exercise, Before/After meal), and press Save. Now press the Back button to exit this screen.

🕑 Note

After a reading is saved into history, it cannot be modified.

If you would like to enter another BG result, exit by pressing the Back button and re-enter the Enter BG Reading Manually screen.

🕑 Note

If the Bolus Calculator is enabled in the Settings Menu, the button will turn green after a blood glucose reading has been recorded. See Section 11

Image: matrix and state in the state in
A
- 6.8 +
mmol/L
Sefore Exercise Sefore Meal
After Exercise After Meal
Save
Bolus Calculator

Figure 101

BG Reminder Settings screen (Figure 102).





10.10 Settings

From the Home Screen, choose Settings/BG Settings. Press BG Reminder Settings to enter the Press 🛨 to add a new reminder and a window will open (Figure 103). Enter the time and the Reminder Name, and press OK.



If you would like for this reminder to repeat every day, click the check box labeled "Repeat this Reminder?"

Pressing the OK Button will return you to the main BG Reminders screen and your new BG reminder will appear. Now, you can enable/disable the reminder by checking/unchecking the box (Figure 104).



Figure 104

🕑 Note

If a BG Reminder is activated, the 💋 icon will appear in the Status Bar.

10.11 Blood Glucose Meter Troubleshooting



BG Alerts consist of one short tone (beep), one short vibration and sometimes a pop-up window that occur simultaneously.

BG Alert Message	Message Type	Solution
BG Meter initialization error	Audio and Vibration Alert with Message Window	Restart your PDA. If the problem persists, contact your distributor.
Test strip was removed during the test	Audio and Vibration Alert with Message Window	Repeat the test and ensure test strip remains in place.
Test strip is contaminated, used, or the blood sample is added to the test strip prematurely	Audio and Vibration Alert with Message Window	Retest with a new strip.
Insufficient sample	Audio and Vibration Alert with Message Window	Retest with a new strip. Make sure there is enough blood to fill the test window.
Temperature exceeds operation range	Audio and Vibration Alert with Message Window	Move to a place within the normal operating temperature range and repeat the test.
Test result is below the measurement range	Audio and Vibration Alert with Message Window	Repeat test. If you see the same result, contact your healthcare professional immediately.
Test result is above the measurement range	Audio and Vibration Alert with Message Window	Repeat test. If you see the same result, contact your healthcare professional immediately.
Check Ketones	Audio and Vibration Alert with Message Window	Check Ketones and contact your healthcare professional immediately.

11 Bolus Calculator

11.1 Introduction

Before administering a bolus, patients usually need to calculate the amount of insulin to administer based on parameters such as the amount of carbohydrates they are eating. This process can be prone to errors because there are many parameters to consider.

The PDA contains a bolus calculator that can suggest a bolus size based on your input. Once the settings are configured properly, the bolus calculator can make a bolus size suggestion after you test your blood glucose level and enter the amount of carbohydrates that you will be eating. The bolus calculator can also take into account the amount of insulin that is currently being used in your body and make corrections to the suggested bolus size.

🙂 Note

The Bolus Calculator requires the following information:

- 1) Current blood glucose level (from internal BG meter)
- 2) Target blood glucose level (user setting confirmed by your healthcare professional)
- 3) Carbohydrate ratio (confirmed by your healthcare professional)
- 4) Carbohydrate intake (user input)
- 5) Insulin sensitivity factor (confirmed by your healthcare professional.)
- 6) Negative/Reverse correction (user setting confirmed by your healthcare professional)
- 7) Active insulin time (confirmed by your healthcare professional.)

The Bolus Calculator suggestion is calculated as follows:

Bolus Suggestion =

Meal Bolus + Correction Bolus - Active Insulin

Meal bolus is used to compensate for increased blood glucose due to eating:

 $Meal \ bolus \ (U) = \frac{Carb \ Intake \ (g)}{Carb \ Ratio}$

Correction Bolus is used to bring your current BG level to the target BG level:

Correction Bolus (U) =

Current BG Level-Target BG Level (mmol/L) Insulin Sensitivity Factor (mmol/L/U)

🕑 Note

If Negative Correction is ON, a correction bolus is always calculated. If Negative Correction is OFF, the correction bolus will only be calculated if your current BG level is HIGHER than the target BG level.

Active Insulin: Insulin is normally absorbed in the body within about 4-6 hours. If you have recently taken a bolus, there could still be active insulin in your body. The Bolus Calculator automatically subtracts the amount of active insulin based on your last bolus and the Active Insulin Time you enter in the settings.

🙂 Note

The Bolus Calculator is disabled by default. Please enable this feature in the Settings menu as described in Section 11.3.

🕑 Note

The Bolus Calculator uses settings that are preset by the user. These settings should be entered under the supervision of your healthcare professional.

11.2 Using the Bolus Calculator

1 The Bolus Calculator function will be enabled after you use the internal blood glucose meter or you enter your BG level manually and save. Press the Bolus Calculator button and the calculator will ask if you will be eating now (Figure 105).



Figure 105

a) If you do not intend to eat and do not want a meal bolus, choose No and skip to step 3.

🕑 Note

If you are not eating, the bolus calculator will not take into account carbohydrates that you will eat. b) If you will be eating, choose Yes and continue to the next step.

2 In the dialog box, enter the total amount of carbohydrates that you intend to eat (Figure 106) and press OK to move to the next step.

🙂 Note

You can estimate the amount of carbohydrates contained in your food by reading the food packaging, consulting carbohydrate counting books or phone apps or by referring to the PDA's food library. The amount of food and preparation method can have a large effect on the total amount of carbohydrates.



3 The next page is the suggested bolus screen (Figure 107). At the top of the screen is the suggested bolus size, the middle of the screen contains the calculation parameters and the bottom of the screen are action buttons.



Figure 107

a) **Bolus Amount:** The suggested bolus amount appears automatically in the bolus

amount area. Pressing the number will open a dialog box that will allow you to change the number.

b) **Calculation Parameters:** This area shows the parameters that have been used to calculate the suggested bolus, including BG Reading, carb intake and active insulin. The last item is the Suggested Bolus amount.

🕑 Note

If the user has enabled the extend bolus feature, these parameters will also be shown.

c) Action Buttons:

- **Extend Bolus:** Press to choose extended bolus features (see Section 6.4).
- **Start:** Press to begin the bolus. A confirmation window will open, press OK to confirm or Cancel to abort.
- **Details:** Press to open a window that describes how the suggested bolus was calculated in detail.

🕑 Note

The Bolus Calculator calculates a suggested bolus value. Please discuss this with your healthcare professional before using this function.

🕑 Note

Your blood glucose reading is only valid for 10 minutes. If you do not issue a bolus within 10 minutes of your test, please test again to calculate a new bolus value.

🕑 Note

If your blood glucose reading is above or below the measurement range, the Bolus Calculator will be disabled.

11.3 Settings

From the Home Screen, you can navigate to the Bolus Calculator settings by pressing Settings then Insulin Delivery Settings.

🙂 Note

The next five settings will only be displayed if the Bolus Calculator function is ON.

Target Blood Glucose Range

Set the default Target Blood Glucose Range first. If desired, you can set different Target Ranges for different time periods.

2 Carbohydrate Ratio

Set the default Carbohydrate Ratio first. If desired, you can set different Carbohydrate Ratios for different time periods.

🕑 Note

Carbohydrate Ratio is defined as the amount of carbohydrates that can be broken down per unit of insulin and is used to calculate the meal boluses. Since everyone has a different metabolism, set this value under the guidance of your healthcare professional.

③ Insulin Sensitivity Factor (ISF)

Set the default Insulin Sensitivity Factor first. If desired, you can set different ISF's for different time periods.

🕑 Note

Since everyones values are different, choose your settings under the guidance of your healthcare professional.

4 Negative Correction

Click to enable or disable the Negative Correction feature.

G Active Insulin Time

Set the default Active Insulin Time desired.

12 History

12.1 Viewing Your History

From the Home Screen, press the History button to view your records.

12.1.1 Daily Log

After you press the History button, the PDA will display the Daily Log if the PDA is held in portrait orientation (Figure 108). In this screen, you can browse your historical records and view daily totals and other statistics.



Figure 108

Press the date near the top of the screen to open a window in which you can choose any calendar day, or press the \checkmark or \triangleright button to scroll to the previous or next day. Below the date, you can see the BG average, Total Carbs, Total Insulin and bolus % of total insulin for the day. The grey area shows time-stamped event entries such as BG readings, changes in Basal Rate and Bolus amounts. You can use your finger to scroll up and down through the list. You may also filter the results by using the check boxes beneath the list.

12.1.2 Graph Display

If you hold the PDA in landscape orientation, daily information will be displayed in a Graph Display mode (Figure 109).



Figure 109

The Graphs show two types of information:

• Blood Glucose/Carbohydrates: For each blood glucose test reading, a circle is shown on the upper graph. Each record of carbohydrate intake is shown with a green vertical line.

Insulin Delivery Amount: Basal rates are shown as the blue waveform. Boluses are shown as blue vertical lines.

Users can display or hide information in a graph by using the checkboxes. Detailed information appears by placing a finger on the graph in the desired area.

12.2 Historical Averages

From the Daily Log screen, press the AVG button to enter the Historical Averages screen.

12.2.1 Historical Averages List

Hold the PDA in portrait orientation, and the screen displays the Historical Averages List (Figure 110). Average blood glucose and insulin delivery information can be seen on this page. Use the top bar to change the number of days to average.



At the bottom of the screen, you can choose different filters to show, for example, the averages of all readings that were taken Before/After Exercise or Before/After a meal.

12.2.2 Historical Averages Calendar

After entering the Historical Averages page, you can view the Calendar mode by rotating the PDA to landscape orientation (Figure 111). This view allows you to see your hourly data in a calendar format. This information includes BG readings, carbohydrates, and boluses.





The left side of the screen shows the time, while the top of the screen shows the day. You can scroll through and tap different time periods to view detailed information regarding BG level, Carbs entered and bolus amounts. At the very top of the screen, you can change the week by pressing the <a> and <a> buttons.

13 General Settings

From the Home Screen, you can navigate to General Settings by pressing Settings and then General Settings (Figure 112).



Figure 112

13.1 Time and Date

a) Press the Date and Time option to open the Date and Time settings.

b) Press the Date option to adjust the date.

c) Press the Time option to adjust the time.

d) 24-Hour Format: Check the box to have the time displayed in 24 hour format, or leave the box empty for 12 hour format.

\Lambda Warning

Execution of the basal rate time periods and history records are directly affected by the time and date settings. Therapy must be suspended before time and date settings can be modified.

13.2 About the System

From General Settings, choose About System.

Software version: Shows the software version of PDA and Pump

PDA Serial Number: Shows the serial number of the PDA.

S Pump Serial Number: Shows the serial number of the pump that is currently being controlled by the PDA. (This number is also printed on the patch pump housing)

13.3 Language

From General Settings, choose Language. The language can be changed using this option.

13.4 Memory Card

From the General Settings screen, choose SD Card.

1 Total Space: Displays the total capacity of your memory card.

Available Space: Displays how much memory is available for storage in your memory card. **History Export:** Exports the history to a tab delimited file onto the memory card for backup.

13.5 Display

From the General Settings screen, press the Display option.

 Brightness: Choose the Brightness option to adjust the display brightness. You can manually adjust the screen brightness, or choose Automatic Brightness to have the PDA adjust automatically.

Screen Timeout: Choose the Screen Timeout option to adjust the amount of time that it takes the display to switch off due to inactivity.

13.6 User Settings

Press User Settings to enter the User Settings.

1) Username: Add the user's name here.

Password: For security, choose "password" to force the user to enter a password during power-on and wake up. The password must be 6 characters, which can be any combination from the letter "A-F" and the number "0-9". If you forget your password, you can use the Pump Serial number to unlock.

Change Password: Use to reset the password.

Restore Factory Settings: All settings reset to the factory defaults.

🕑 Note

Once the factory settings have been restored, all saved settings will be lost (except for time). Please make a note of all important settings before restoring factory settings.

13.7 Bluetooth

🕑 Note

The Bluetooth function requires a microSD card to be installed.

From the General Settings screen, choose Bluetooth (Figure 113).



Figure 113

1 Bluetooth On/Off: Check the checkbox to switch the Bluetooth on. The Bluetooth logo will appear in the status bar. More options will appear if Bluetooth is on.

2 Machine Name: Choose Machine Name to give your PDA a unique name.



Figure 114

3 Visibility: Check the checkbox to allow other Bluetooth devices to detect the PDA. This function will timeout after 120 seconds.

4 Search for Devices: Press this button to search for other Bluetooth devices that are within range. A list of devices will appear in the Bluetooth Devices section (Figure 115). The status of the

device (paired, unpaired) will be shown underneath the device name.





5 Using the devices on the list, the PDA has the following Bluetooth options:

If a device on the list is not paired, press the device name to pair it with the PDA and a pop up window will open to enter the pairing code. Enter the code and accept to pair a new device (Figure 116).

Municoth	.
BLUETO	отн
Pair with MicroTeo	ch Medical
ок	Cancel
Parsene Carpon	

Figure 116

• If a device is already paired with the PDA, pressing the device name will open a dialog box (Figure 117) that will allow you to either export history via Bluetooth or unpair the device.





14 Additional Features

14.1 Audio Player

The PDA also contains an audio player that can be used to play training instruction audio files. From the Home Screen, choose the Actions button and then the Audio Player button.

14.2 Auto-Off

The pump system also has an automatic shutdown (auto-off) feature. This function can be enabled by checking the check box on the Settings - Insulin Delivery Settings -Auto Off and then choosing the time delay before automatic shut down.

When the auto-off feature is switched on, the pump will automatically stop delivery if there is no user input (button push) for a set amount of time. 15 minutes before shutdown, the PDA will begin to beep to issue a warning that auto-off will occur soon. If there is still no button push, the pump will stop delivery and both the pump and PDA will issue an alarm to tell the user that delivery has stopped.

🕑 Note

The auto-off feature is disabled by default. If the pump ceases operation because of the auto-off feature, the reservoir will need to be replaced to continue operation.

14.3 Food Database

The food database is an extra function that can provide carbohydrate content data for various foods.

From the Home Screen, choose the Actions button and then Food Database. Not only can you browse the database, you can manually add, edit, and delete food data.

15 Suspend/Resume

15.1 How to Suspend/Resume

Occasionally, you may need to temporarily stop insulin delivery by using the Suspend function. This can happen, for instance, if you temporarily do not require any insulin infusion, or need to remove the pump from the infusion set.

From the Home Screen, choose the Suspend button to temporarily stop insulin delivery.A confirmation dialog box opens, as shown in Figure 118, click OK to suspend or Cancel.

You can also choose to stop delivery and allow the pump to rewind. Remember that rewinding the pump will require priming the reservoir again.





After Suspend mode has been activated, the PDA returns to the Home Screen, and the information area of the Home Screen will display the amount of time that insulin delivery has stopped for. The Pause button will become a Resume button, as shown in Figure 119.

🕑 Note

While in Suspend mode, the PDA will beep every 15 minutes to remind you that delivery has stopped.



Figure 119

If you would like to resume insulin delivery, press the Continue button. A confirmation dialog box will open (Figure 120). Choose OK to resume insulin delivery.





Note

If a bolus is in progress, you cannot enter Suspend mode. You must first cancel the bolus before suspending delivery.

16 Alarms and Troubleshooting

The patch pump system has a comprehensive safety system to check if abnormal situations require immediate attention. The system will send notification alarms using sound, LEDs, or vibrations as well as provide information on the PDA display.

The insulin pump system contains two parts – the patch pump detects and sends the alarms, and the PDA receives the alarms and notifies the user.

The insulin pump system only contains medium and low priority alarms – no high priority alarms (as defined by ISO standards).

Medium priority alarms occur when the delivery function has stopped due to technical failure, requiring the user to intervene in the operation of the pump, change the pump, or possibly inject insulin manually to control BG levels. Low priority alarms alert the user that something will occur in a short period of time but insulin delivery continues normally. Users should be aware of this information and make plans in advance to ensure that treatment continues reliably.

🕑 Note

There are no alarms to indicate old insulin or to detect insulin leaks. The user should be aware if these situations happen.

🕑 Note

If the insulin pump and PDA start up normally without alarms, the alarms system is working properly.

🕑 Note

You cannot modify the alarm settings, including alarm volume.
PDA Alarm Priority Level s

Alarm Level	Visual Signal	Audio Signal
Medium Priority	Flashing Yellow Light	Three consecutive beeps
Low Priority	Steady Yellow Light	Two consecutive beeps

Patch Pump Alarm Priority Levels

Alarm Level	Visual Signal	Audio Signal	Vibration Signal
Medium Priority	Flashing Yellow Light	None	Yes
Low Priority	Steady Yellow Light	None	Yes

Audio Alarm Volume

Device	Sound Pressure Level (dB)	
Patch Pump	None	
PDA	60-90	

16.1 Patch Pump Alarms

Alarm Description	Priority Level	Alarm Signal	Solution/Action
No Insulin Remaining	Medium	Vibration	There is no more insulin and delivery has stopped. The pump will automatically rewind. Replace the reservoir and check your blood glucose level.
Blockage Detected (Occlusion)	Medium	LED, Vibration	Insulin delivery has stopped. Replace the reservoir/infusion set and check your blood glucose level.
Unexpected Delivery Stoppage	Medium	LED, Vibration	Insulin delivery has stopped. Check your BG level. Replace reservoir/infusion set. If problem persists, contact the supplier for guidance.
Pump Battery Exhausted	Medium	LED, Vibration	Insulin delivery has stopped. The pump will automatically rewind. Please replace the depleted battery with a fully charged one.
Cannot Give Full Bolus/Dose Low	Low	LED, Vibration	There is not enough insulin left in the reservoir to execute the requested bolus. Prepare a new reservoir to be used after the old one is empty.
Pump Battery Level is Low	Low	LED, Vibration	The pump battery level is below 5%. Please prepare a fully charged battery for replacement

🕑 Note

"Unexpected delivery stoppage" When this notice appears, the pump infusion has been stopped and the pump will automatically rewind. Please check the battery connection and the pump status, Measure your blood glucose level.

🕑 Note

If the "Auto-Off" function is enabled, infusion may automatically stop when preprogrammed this way. For details, please refer to 14.2.

16.2 PDA Alarms

Alarm Description	Priority	Alarm Signal	Solution/Action
No Insulin Remaining	Medium	LED, Vibration	The reservoir is empty and insulin delivery has stopped. The patch pump will automatically rewind. Replace the reservoir and check your blood glucose level.
Blockage Detected (Occlusion)	Medium	LED, Vibration	Insulin delivery has stopped. Replace the reservoir/ infusion set and check your blood glucose level.
Unexpected Delivery Stoppage	Medium	LED, Vibration	Insulin delivery has stopped. Check your BG level. Replace reservoir/infusion set. If problem persists, contact the supplier for guidance.
Pump Battery Exhausted	Medium	LED, Vibration	Delivery has stopped. The patch pump will automatically rewind. Please replace the depleted battery with a fully charged battery.
Cannot Give Full Bolus/Dose Low	Low	LED, Vibration	There is not enough insulin left in the reservoir to execute the requested bolus. Prepare a new reservoir to be used after the old one is empty.
Pump Battery Level is Low	Low	LED, Vibration	The pump battery level is below 5%. Please prepare a fully charged battery for replacement
PDA battery exhausted	Low	LED, Vibration	The PDA battery level is below 5%. Connect to charger.
PDA Failure	Low	LED, Vibration	Restart the PDA. If the problem persists, contact your supplier for a replacement.

•

If a medium priority alarm occurs, the patch pump will automatically rewind and stop all insulin delivery. A pop-up window will appear on the PDA display to show more information and proposed solutions to the alarm. Press "OK" to close the pop-up window. Replace the reservoir to re-initialize the system and continue delivery.

🕑 Note

When the PDA gives an alarm signal, it will also vibrate to ensure the user is notified.

🕑 Note

Low priority alarms appear only once and do not repeat

16.3 Alarm System Delay

The alarm system has an inherent delay between the two components, as shown in Figure 121:

- T2-T1: The time it takes for a safety sensor in the patch pump to detect an alarm ≤0.1s.
- T3-T2: The amount of time between detecting an alarm and issuing an alarm signal within the patch pump ≤0.1s.
- T4-T3: The time it takes for the pump to send the alarm information wirelessly to the PDA ≤ 4s.
- T5-T4: The amount of time between when the PDA receives the alarm to issuing the alarm signal ≤0.1s.



Graphical Representation of Alarm Delay

The alarm system is distributed – if a pump sensor detects a problem it will issue an alarm within 0.2 seconds. The PDA will receive the information and issue an alarm within 4 seconds.

17 Maintenance

Your patch pump and PDA are precision instruments. Improper use and maintenance will result in decreased accuracy or even pump failure. Please read this section carefully to learn how to properly care for your patch pump system.

17.1 Cleaning

17.1.1 Pump

1 Clean the outer surface using a mild detergent and a soft damp cloth. Use another cloth to dry.

- 2 Disinfect with an alcohol wipe.
- **3** Do not use solvents, nail polish remover or paint thinner to scrub the outer surface.
- 4 Keep the pump dry, avoid water.
- 5 Do not use any lubricant.

17.1.2 PDA

1 Clean the outer surface using a mild detergent and a soft damp cloth. Use another cloth to dry.

- 2 Disinfect with an alcohol wipe.
- **3** Do not use solvents, nail polish remover, or paint thinner to scrub the outer surface.
- 4 Keep the PDA dry, avoid water.
- 5 Do not use any lubricant
- 6 Keep the test strip port area clean.

17.1.3 Lancing Device

As needed, use a soft damp cloth moistened with soap and hot water to wipe the surface. Do not immerse the lancing device in water.

17.2 Avoid Extreme Temperatures

- 1 Avoid exposing the pump and PDA to temperatures above 40°C and below 0°C.
- 2 Insulin solutions freeze near 0°C and degrade at high temperatures. If you

are outside in cold weather, wear your pump close to your body and cover it with warm clothing. If you are in a warm environment, take measures to keep your pump and insulin cool.

3 Do not steam sterilize or autoclave your pump or PDA.

17.3 Avoid Immersion in Water

The patch pump is rated IPX4 (splashproof). Do not bathe, swim, or otherwise immerse your pump in water. If you accidentally soak the pump with water, it may not function correctly.

The PDA is not waterproof, and thus should be kept away from water at all times.

17.4 Test Strips

• Use only MicroTech's Exactive EQ glucose test strips with the PDA.

• Store test strips in a clean, dry environment at 5-30°C (41-86 °F). Do not

store test strips in heat or direct sunlight.

- Do not refrigerate or freeze test strips.
- Do not store or use strips in a humid environment, such as a bathroom.
- Do not store the PDA, test strips or control solution near bleach or cleaning agents that contain bleach.
- Close the cap on the vial immediately after removing a test strip.
- Use the test strip immediately after removing it from the package.
- Do not use expired test strips. Doing so may lead to inaccurate results.

🕑 Note

The test strip label contains the expiration date in year-month format. For example, 2019-01 indicates that the test strips are valid until January 2019.

Special Instructions for Test Strips Sold in a Vial:

• Test strips should be stored in the tightly capped vial that is provided.

• Do not store test strips outside of the provided vial. Test strips must be stored inside the original vial with the lid tightly closed.

- Do not transfer test strips from the provided vial into another container.
- Close the cap on the vial immediately after removing a test strip.

• A new vial of test strips may be used for 6 months after first being opened. Please take note of the date that the vial was first opened and discard after 6 months.

Test Strip Precautions:

For in vitro diagnostic use.

- Use the test strip immediately after removing it from the package, otherwise the test results may not be accurate.
- Do not use test strips that are torn, bent or damaged in any way. Do not reuse test strips.

- Keep the test strip packaging away from children and pets.
- Consult your physician or healthcare professional before making any changes in your treatment plan based on your blood glucose test results.
- Please refer to the test strip instructions for more detailed information.

17.5 Control Solution

Control solution is a glucose solution of known concentration that is used to confirm that your PDA blood glucose meter and test strips are working properly. It is important to run a quality control test regularly to make sure that you are getting accurate results.

You should perform a quality control test in the following situations:

- When you suspect that the meter or test strips are not working properly.
- When you suspect that your test results

are inaccurate or inconsistent with how you feel.

- When you suspect that your meter has been damaged.
- After cleaning your PDA.

Refer to Section 10.8 for instructions on how to perform a quality control test.



Storage and Handling

Please review the following storage and handling instructions:

- Store the control solution in the temperature range 5-30°C (41-86°F).
- Do not refrigerate or freeze the control solution.

- If the control solution is cold, do not use until it has warmed to room temperature.
- Do not use expired control solution.

🙂 Note

The control solution label contains the expiration date in year-month format. For example, 2019-01 indicates that the control solution is valid until January 2019.

• Control solution may be used for 6 months after the bottle is opened for the first time. Please take note of the date that the bottle was first opened and discard after 6 months. Do not use beyond the expiration date.

Control Solution Precautions:

- For in vitro diagnostic use. The control solution is for testing only outside of the body. Do not swallow or inject.
- Control solution should be shaken before use.
- Quality control tests should be carried

out at 15-30°C.

- Do not let the control solution bottle touch the test strip.
- Use only the control solution that is recommended for your meter.
- The control ranges shown on the test strip package are not recommended ranges for your blood glucose level. Your personal glucose range should be determined by your healthcare professional.

Please refer to the control solution instructions for more detailed information.

17.6 X-Ray, MRI and CT Scans

If you are going to have an X-ray, CT scan, MRI or other type of exposure to radiation, remove your pump and PDA before entering the room that contains this equipment.

17.7 Precautions

Although the pump has multiple safety alarms, it cannot notify you if the infusion set is leaking or if the insulin has lost its potency. It is essential, therefore, that you test your blood glucose levels at least four times per day. If your blood glucose is out of range, check the pump and infusion set to ensure that the necessary amount of insulin is being delivered.

17.8 Wireless Connection

The patch pump and PDA communicate wirelessly, and when the PDA sends instructions to the pump, they must be within an acceptable distance. The patch pump and PDA have a wireless communication range of 2 meters. The distance and surrounding environment have a large effect on the wireless signal integrity.

Please follow the suggestions below to

maximize the wireless signal.

1 Avoid obstructions between the PDA and pump, such as walls, floors, metal plates, people, etc.

2 Avoid wearing clothing that contains metallic substances around the pump.

6 Avoid strong electromagnetic radiation.

4 Keep other wireless devices away from the pump and PDA, even if the devices comply with national emission requirements. Wireless interference can still occur.

If the signal strength between the patch pump and PDA is good, information will travel more quickly between the two. Always observe the wireless signal strength on the status bar before using the PDA. If the signal is weak or there is no signal, the PDA will not be able communicate with the pump.

🙂 Note

If the signal is weak or there is no signal, check that you are avoiding the above four situations. If the signal is still weak, move the PDA closer to the pump. If the situation persists, please contact your supplier.

17.9 Disposal

When you replace the pump, PDA and its attachments, please do not throw them away. Take them to an electronics recycling facility or return to your supplier.

Do not discard damaged or expired batteries. Please dispose of batteries according to local recycling laws.

17.10 Transportation

Avoid placing heavy objects on top of the patch pump and PDA. Avoid direct sunlight and rain. Transport according to the transportation conditions.

17.11 Storage

If you are temporarily not using the pump system, store the components in a cool, dry, clean and well-ventilated area.

If you decide not to use the pump for a prolonged period, the battery should be stored separately.

17.12 Other Considerations

When dealing with potentially contagious substances (such as blood or reagents), use protective gloves or other protective coverings if skin exposure is likely.

18 Specifications

18.1 General Specifications

	Patch Pump	PDA	
Model Number	MTM-1	MTM-2	
Size	59.5x40x11.1mm (LxWxD)	112x57.2x12mm (LxWxD)	
Weight	23g (without batter or insulin)	71g (without battery)	
Reservoir Capacity	2ml	-	
Operating Temperature	5-40°C (41-104)	5-40°C (41-104)	
Operating Humidity	10-93% (noncondensing)	10-93% (noncondensing)	
Storage Temperature	-40°C-55°C	-40°C-55°C	
Storage Humidity	5-95% (noncondensing)	5-95% (noncondensing)	
Water Resistance	IPX4	IPX0	
Alarm Prompts	LED (Yellow) , Vibration	Audio, LED (Yellow), Screen	
History Storage	Automatically Syncs to PDA	Browse by screen	
Display	None	3.2 inch color touchscreen	
Battery	70mAh 1000mAh		
Low Reservoir Alarm Threshold	10-50U, in increments of 5U	, 10U by default	
Auto-Off	Enable/Disable – Disabled b	y default	
Auto-Off Time Range	1-24 hours, 1 hour incremen	ts, 10 hours by default	
Memory Behavior During Power Off	All settings and records retained after power off.		
Wireless Frequency and Bandwidth	Frequency: 2.402GHz~2.48 GHz Bandwidth: 1Mbps		
Wireless Modulation	GFSK		
Radiated Power	-2dBm		
Warranty	4 Years		

18.2 Delivery	
Attribute	Specification
Basal Rate	0.025 - 35 U/hr, programmed in increments of 0.025U/hr
Basal Programs	3 Basal Programs, each with 48 time segments
Maximum Basal Rate	0.1-35 U/hr, Default: 1.5 U/hr
Base Basal Rate	0.025-35 U/Hr, Default: 0.5 U/hr
Temporary Basal Rate	U/hr or $\%$ of basal rate, last 3 basal rates remembered, Default is Off
Bolus Size	0.025U-25U, 3 Presets
Bolus Increment	0.025/0.05/0.1/0.5/1 U, Default:0.1 U
Max Bolus Size	1-25U, programmed in 1U increments, Default is 10U
Extended Bolus	Programmed in U or % of total bolus,Default is Off
	Extend time: 0.5-8 hours in increments of 0.5 hr
Quick Bolus	On/Off, default is off
Quick Bolus Increment	0.1-2U, Default is 0.1 U

18.3 Blood Glucose Meter

Attribute	Specification
Measurement Range	1.1-33.3 mmol/L
Test time	5 Seconds
Test Reminders	7 Reminders, which can be repeated
HCT Range	30-55%

18.4 Bolus Calculator

Attribute	Specification
Bolus Calculator	On/Off, default is off
Carb Ratio	1-150 g carb/U in 1 g carb/U increments, no default
Insulin Sensitivity Factor	0.1-16.7 mmol/L/U in 0.1 steps, no default value
Negative Correction	On/Off, Default is On
Active Insulin Time	2-6 hrs in 0.5hr increments, no default value

18.5 Bolus Delivery

Bolus Steps	Volume	Time Interval	Infusion Speed
	per Step	Between Steps	per Minute
0.05 U	0.5 μL	1s	3 U

18.6 Infusion Precision

At 1U/h delivery rate, the error was measured at 0.4% as shown in Figure 122.

🕑 Note

The above test results were achieved using pump serial number 0A001 and reservoir batch number G13B25001





When the delivery rate was set to 0.01mL/hr, the measured flow is shown in Figure 123.



Figure 123

18.7 Occlusion Detection (Maximum Infusion Pressure)

When the pressure inside of the reservoir reaches a maximum of 100kPa±30kPa, the occlusion alarm will occur and the motor system will rewind automatically.

18.8 Occlusion Alarm Time

When a fluid obstruction is detected, an occlusion alarm occurs. An average of 2.5U of insulin will be delivered before this alarm appears.

The Table below shows two delivery speeds and the respective occlusion alarm delays using U100 insulin.

Rate	Typical Time Before Alarm	Maximum Time Before Alarm	
Fast Rate (3U/hr)	50 sec	53 sec	
Medium Rate (1U/hr)	150 min	160 min	

18.9 Overdose/Underdose

The pump contains sensors that verify the infusion accuracy.

If the actual amount of delivery is more or less than the requested amount, this is called an overdose or underdose. The sensors in the pump can detect overdose or underdose situations and compensate automatically, or issue an alarm.

The maximum amount of bolus that can be delivered in a single fault is 0.25U.

18.10 Electromagnetic Compatibility

These devices are intended for use in the electromagnetic environment specified in this section. The user of the device should ensure that the device is used in such an environment.

Portable and mobile RF communication interference may have an impact on the device.

Please use the cables and accessories provided. The cable information is as follows:

#	Item	Length (m)	Cables Shielded	Notes
1	PDA Charge Cable	1.0m	Yes	EUT DC 5V

The use of accessories other than those specified for the device is not recommended. They may result in increased emissions or decreased immunity of the device.

The device should not be used adjacent to or stacked with other equipment. If adjacent or stacked use is necessary, the device should be observed to verify normal operation in the configuration in which it will be used.

The basic performance is described in the table below:

Performance	Specific Description	

Infusion Precision within the $\pm 5\%$

Guidance and manufacturer's declaration - electromagnetic immunity

The device is intended for use in the electromagnetic environment specified below. The user of the device should ensure that it is used in such an environment.

Emissions Test	Compliance	Electromagnetic environment - guidance
RF Emissions CISPR 11	Group 1	The device uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF Emissions CISPR 11	Class B	The device is suitable for use in all establishments, including domestic establishments and those
Harmonic Emissions IEC 61000-3-2	Class A	directly connected to the public low-voltage power supply.
Voltage Fluctuations/Flicker Emissions IEC 61000-3-3	Complies	

IEC 60601-1-2: Table 202

Guidance and manufacturer's declaration – electromagnetic immunity

The device is intended for use in the electromagnetic environment specified below. The user of the device should ensure that it is used in such an environment.

Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic environment – guidance
Electrostatic discharge (ESD) IEC 60601-4-2	±6 kV Contact ±8 kV Air	±8 kV Contact ±15 kV Air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient burst IEC 61000-4-4	±2 kV ±1 kV for input/ output lines	±2 kV ±1 kV for input/ output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	$\pm 1~\text{kV}$ Differential mode $\pm 2~\text{kV}$ Common mode	± 1 kV Differential mode ± 2 kV Common mode	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11	<5 % Ut >95 % dip in Ut) for 0.5 cycle 40 % Ut (60 % dip in Ut) for 5 cycles 70 % Ut (30 % dip in Ut) for 25 cycles <5 % Ut (>95 % dip in Ut) for 5 s	<5 % Ut (>95 % dip in Ut) for 0.5 cycle 40 % Ut (60 % dip in Ut) for 5 cycles 70 % Ut (30 % dip in Ut) for 25 cycles <5 % Ut (>95 % dip in Ut) for 5 s	Mains power quality should be that of a typical commercial or hospital environment. If the user of the device requires continued operation during power mains interruptions, it is recommended that the device be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	3 A/m	400 A/m	The power frequency magnetic field should have the characteristics of power frequency magnetic field level at a typical place in a typical commercial and hospital environment

IEC 60601-1-2: Table 204

Guidance and manufacturer's declaration - electromagnetic immunity

The device is intended for use in the electromagnetic environment specified below. The user of the device should ensure that it is used in such an environment.

Immunity Test	IEC 60601 Test Level	Compliance Level	Electromagnetic environment – guidance	
Conducted RF IEC 61000-4-6	3V (Vrms) 150kHz~80MHz 10V (Engineering medical frequency band) 150kHz~80MHz	3V (Vrms) 10V (Engineering medical frequency band)	Portable and mobile RF communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance d=1.2 √P d=1.2 √P 80MHz~800MHz d=2.3 √P 800MHz~2.5GHz where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m).Field strengths from fixed RF	
Radiated RF IEC 61000-4-3	10 V/m 80MHz~2.5GHz	3 V/m	transmitters, as determined by an electromagners site survey, ^a should be less than the compliance in each frequency range. ^b Interference may occur in the vicinity of equipm marked with the following symbol:	

- a Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the device is used exceeds the applicable RF compliance level above, the device should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the device.
- **b** Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

Recommended separation distances between portable and mobile RF communications equipment and the device

These devices are intended for use in an environment in which radiated RF disturbances are controlled. The user of the device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the device as recommended below, according to the maximum output power of the communications equipment

Maximum rated	Separation distance according to frequency of transmitter (m)			
output power of transmitter	150kHz~80MHz d=1.2 √P	80MHz~800MHz d=1.2 √P	800MHz~2.5GHz d=2.3 √P	
0.01	0.12	0.12	0.23	
0.1	0.38	0.38	0.73	
1	1.2	1.2	2.3	
10	3.8	3.8	7.3	
100	12	12	23	

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

19 Appendix

19.1 Symbols

Single Use Only	\otimes	Temperature Range	X
Consult Instructions for Use	i	Handle With Care	Ţ
Use By	2	Type BF Applied Part	Ŕ
Manufacturer		Water Resistance Level	IPX4
Lot Number	LOT	Recyclable	
Serial Number	SN	Do Not Dispose with Household Waste	Z
Sterilized by EO	STERILE EO	Keep Dry	Ť
Biohazard	ගිව	Control Solution Range	CTRL
In Vitro Diagnostic Device	IVD	Keep Away From Heat and Radiation Sources	~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~
Non-Ionizing Radiation		Class 2 Equipment	
See Instructions for Use	Â		



MicroTech Medical (Hangzhou) Co., Ltd.

Address: No.108 Liuze St., Čangqian,Yuhang District, Hangzhou, 311121 Zhejiang, P.R.China Phone: 011 -86-571-56782339 Website: http://www.microtechmd.com

EC REP Lotus NL B.V.

Koningin Julianaplein 10, 1e Verd, 2595AA, The Hague, Netherlands.

Local distributor:

Homemed (Pty) Ltd

Address: 100 Sovereign Drive, Route 21 Corporate Park, Irene, Centurion, South Africa Phone: 0861 106 150 Website: https://www.homemed.co.za/ 1001-PMTL-256 V1.0

(E 0197